

MIMER Medical College, Talegaon Dabhade
Log Book

Date	Department	Machine Name	Breakdown	Action Taken
03-06-2021	Paediatric OPD	CPU Fan	CPU fan Error	Replace the fan
28-06-2021	Accounts Hospital	Operating system	Continuous restarting	After maintenance & Extra RAM installation machine working
16-07-2021	PG Section	Operating system	Windows update problem	After uninstalling recent update woking
17-08-2021	Purchase	Operating system	Operating system corrupted	Installed fresh copy of Win 10
30-08-2021	Research lab	Operating system	After login System hang	Installed fresh copy of Win 10
02-09-2021	Anatomy	CPU fan	CPU fan Error	Replace the fan
11-09-2021	Pharmacology	Printer	Paper stuck in printer	After removing paper working
23-09-2021	Central Store	CPU	Continuous restarting	Repair with chkdsk /r
11-10-2021	CCL	Printer	Paper stuck in printer	After removing paper working
30-11-2021	Cash Counter	Printer	Paper stuck in printer	After removing paper working
04-01-2022	Medicine	CPU	RAM	Change RAM
06-01-2022	Anaesthesia	CPU	SMPS (Power Supply Problem)	Change SMPS
21-01-2022	Admin hosp	CPU	Motherboard problem	Replace the motherboard
02-02-2022	PSM demo	RAM	Computer not working, red blink	Change RAM
08-02-2022	For College Building	Projector	Panel repair & servicing	Projector are working ok
18-02-2022	Reg counter	CPU USB port	USB not detected	RAM changed
21-02-2022	Patho clerk	Opearting system	Continue restarting	Repair with CD
10-03-2022	Paeds ward Network cable	N computing device	Internet not working	Change the LAN Cable
29-03-2022	Ortho ward cable	N computing device	VGA cable problem	Change the VGA cable

31-03-2022	Purchase Dept	Opearting system	Desktop files deleting	Format CPU with win 10 CD
02-04-2022	Cash counter	Opearting system	Continue restarting	Chkdsk /r
06-04-2022	Admin office	HDD	Not able to open files	Change HDD
21-04-2022	Research lab	CPU Cmos cell	Date & time issue	Change Cmos cell
15-06-2022	ICT	I-card Printer	Servicing of Printer	Working ok after repair
10-08-2022	OT	CPU	SMPS (Power Supply Problem)	Change SMPS
12-09-2022	MRD	Printer	Paper Pick-up Problem	Chnage the Part (Paper Pickup Assembly)

April - 2022

13/04/22 F.M.T. Yoga session for 1st year students on following dates by yoga teacher Ruchira Kulkarni.
Venue: Auditorium, Hospital Building

1st	15/04/22	-	9.00 am to 11.00 am
2nd	16/04/22	-	9.00 am to 11.00 am
3rd	17/04/22	-	11.00 am to 1.00 pm
4th	18/04/22	-	11.00 am to 1.00 pm
5th	19/04/22	-	11.00 am to 1.00 pm
6th	20/04/22	-	11.00 am to 1.00 pm

15/04/22	Key Handover (cleaning work done)	RF
16/04/22	Key Handover (cleaning work done)	RF

11/04/22 OBGY Ward No - B-106. 1) In the Doctors wash room the flush tank is not working and needs to be repaired or replaced

12/04/22 flush repaired work done. Bapu Sakharekar RF

2) In the ward water cooler drainage is block needs to be repaired.

12/04/22 checkup removed work done. Ashok Walmiki RF

14/04/22 M.M.W B-20 patient toilet is checkup and flush is not working.

15/04/22 checkup removed work done. Ashok Walmiki RF

Khandys
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE,
PUNE 410507

RF
Maintenance Dept.
MIMER Medical College
Talegaon, Dabhade

Khandys
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE 410507

RF
Maintenance Dept
MIMER Medical College
Talegaon Dabhade

19/04/22 Student Council
Our sports events for plexus is starting from 21st April. For Indoor games we will be needing
1) for chess - chair - 08 -
Tables - 04
2) for Carrom - stand - 05 Net stool - 20 Nets

20/04/22 Blood Bank
In Blood Bank cross matching room wash basin is not working properly. water leakage is there. So please repair.

20/04/22 Biochem's Dept.
The requirement of fume hood & Basin taps - of practical hall.

20/04/22 Circular
Mrs. Vajalawany prof. & Director Academic, D.Y. Patil for HACC peer team visit.

Khandy
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE,
PUNE - 410507

Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

20/04/22 Sport Event cancel by college.

SLB

26/04/22 New wash basin & tap replaced. plumber work done.

SLB

This complaint forwarded to Artist dept. plumbing work done.

SLB

22/04/22 - provide slushout hall
2.30 pm - 3.30 pm Feeding staff
3.30 pm - 4.00 pm Non Feeding staff.
(cleaning work done.)

SLB

Khandy
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE - 410507

Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

21/04/22 Peads 1) Pediatrics OPD Room No. 4/
OPD drainage pipe of basin has
been broken. so please need fix

25/04/22 Gynoc - wash basin pipe 03 No. has been
OPD. Broken

21/04/22 Boys Wash Basin taps have been
Hostel Leaker head to be fixed

2/05/22 wash pipe replaced. plumber S.S.

21/05/22 New pipe replaced in 03 No. gynoc
V gynoc dept plumber S.S.

25/05/22 All Basin tap fitting
work done plumber S.S.

Khandays
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE,
PUNE 410507

Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

Khandays
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE 410507

Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

May - 2022

19/05/22	MRT 501 Colleges.	1) Window Glass broken in class room & Wash Rooms and need to replaced before rainy seasons 2) plumbing work repair leakage in washrooms.	3/6/22	New window Glass replaced. Acetic done	Suresh Gud	SL
			23/05/22	plumbing work done	plumber	SL
21/05/22	21/05/22 UHTC. Talegaon	Kindly depute a sweeper to Urban Health Centre to cleaning toilet with comode.	21/05/22	work done	Ashak	SL
21/05/22	21/05/22 Community medicine	1) Kindly replaced 03 glass doors of the book self shelf at Urban health centre.	27/05/22	03 Glass fitting work done	demetri	SL
24/05/22	Boys Hostel	1) Valve of mess water tank is not working properly. need to be replaced	25/05/22	new valve replaced work done	plumbers	SL

Khandays
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE 410507

SL
Maintenance Dept.
MIMER Medical College
Talegaon Dabhaide

Khandays
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE 410507

SL
Maintenance Dept.
MIMER Medical College
Talegaon Dabhaide

26/05/22 ophthalmology Dept 1) O.P.D No. 33, Tab knob is broken please repair.

27/05/22 Tab repair by plumber work done plumber SL

25/05/22 orthopaedic Department 1) A Guest Lecture has been arranged by Dept. ortho on 26/05/22 Time 11.30 am at Sushant Hall.

26/05/22 cleaning work done Hardwar to Hall to ortho Dept. SL

30/05/22 ortho ward water cooler outside make ortho water leakage, water accumulation outside. fan noisy and in the corridor.

7/06/22 water cooler outlet pipe checkup & we repair work done (New pipe fitted) outside of ward. plumber SL

Principal
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE - 410507

SL
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

Principal
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE 410507

SL
Maintenance Dept
MIMER Medical College
Talegaon Dabhade

June - 2022

1/06/22	female surgery ward	1) 02 Tiles broken in female surgery ward.	2/06/22	Tiles fitting work done.	SLB
7/06/22	paeds OPD.	1) water leakage from Basin pipe in paeds OPD.	9/06/22	New Basin Tap replace & pipe	plumber SLB
9/06/22	B-20 Male medicine ward.	1) Tiles in front of counter is broken	10/06/22	work done.	SLB
9/06/22	Boys Hostel	1) All 4 floors - wash Basin inlet pipe need to be changed. 2) All 4 floors - wash Basin outlet pipe need to be changed.	10/06/22	plumbing work done. All inlet pipe replaid & Basin pipe work done.	plumber SLB

Khandys
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE - 410507

SLB
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

Khandys
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE - 410507

SLB
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

14/06/22	Dentistry Xph	1) There is leakage of the pipe of wash basin in dental opD. Kindly repairs as soon as possible.	14/06/22	work done.	Shrin	SL
15/06/22	Pathology Dept.	1) few floors. files in pathology practical hall are broken kindly do the needful urgent basis.	17/06/22	work done.	mehar	SL
16/06/22	Blood Bank.	1) In Blood Bank Autoclave Room Wash Basin is not working properly, water tap leakage. do the repairs.	18/06/22	work done, new tap replace.	plumber.	SL
17/06/22	Biochemistry Dept.	1) please repair distilled water plant fume hood. urgent basis.	18/06/22	distilled water plant leakage repaired. work done.	plumbers	SL
18/06/22	skin ward.	1) Wash room flush is not working properly do the needful.	18/06/22	new flush replaced.	plumbers	SL

Khandys
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE 410507

SL
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

Khandys
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE 410507

SL
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

17/06/22	Optical Dept. (Wash Room)	1) Basin pipe is not working properly, in common wash room. please repair.	18/06/22	Basin pipe is chocked so chocked removed	Ashok	Self
20/06/22	PNC Room-206	1) Patients washroom pipes block that do the needful	21/06/22	work done.	Ashok	Self
25/06/22	B-20 (M.M. Ward)	1) Bathroom of patient toilets is blocked.	28/06/22	Block repaired. work done	Ashok	Self
22/06/22	UHTC Talegaon Community Medicine.	1) kindly repair flush tank of Commode at UHTC.	24/06/22	New flush tank fit in the plunger of UHTC, Talegaon		Self
22/06/22	Optical Dept (Wash Room)	1) water Basin water draining blocked please repair.	22/06/22	work done	Ashok	Self

Khandys
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE,
PUNE 410507

Self
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

Khandys
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE 410507

Self
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

July - 2022

1/07/22 Orthopedics 2) In mate cotno heard there is a leaking of water in kindly do the needful as possible etc.

2/07/22 OBST & Gynaec 1) please change the tap in Dept. SPD - No-46 (207C)

2/07/22 OBST & Gynaec D patient toilet wash basin tap is not working
Ward No. 106.

2) In ward No. 102, staff toilet hand spray, tap not working. please do the needful.

5/7/22 Ward No. 304 - D) water leakage from the ceiling side in ward.

7/07/22 B-22 Ward. D) patient toilet glass window is broken.

K. S. Khandys
PRINCIPAL
MIMER MEDICAL COLLEGE
TALGAON DABHADE
PUNE 410507

Maintenance Dept.
MIMER Medical College
Talgaon Dabhade

1/07/22 construction problem do the complaint given to the department. etc.

5/07/22 Wash Basin tap repaired by plumber etc.

8/07/22 Basin tap replaced in new & new hand spray fitting work done. etc.

6/07/22 This problem of construction dept. Expansion Joint of Building leakage etc.

9/07/22 New Glass of window fitting work done. etc.

K. S. Khandys
PRINCIPAL
MIMER MEDICAL COLLEGE
TALGAON DABHADE
PUNE 410507

Maintenance Dept.
MIMER Medical College
Talgaon Dabhade

21/07/22	Anesthesiology Dept.	1) In operation theater two capheads glass is broken so please look it	21/07/22	work done.	SL
21/07/22	CCL.	In cct scatology section Resin Block so please do the need full	21/07/22	block forward work ok	SL
21/07/22	Surfery Dept.	Guest Lecture Date - 21/07/22 100 pm in auditor Hall	21/07/22	Hall Handover to Surfery dept.	SL
21/07/22	Ward No. 106.	Patient washroom Tap is not working and need to replace	21/07/22	work done. New pipe fitting	SL
21/07/22	Blood Bank	Blood Donation Camp is fixed in G.J. Ghaisas Hall OPD on 22/07/22 Material Required - Tabled Fan - 08 plastic chair - 40 Wooden table - 05	22/07/22	work done. Hall cleaning done.	SL

K. S. D. S.
Principal
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE - 410507

SL
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

K. S. D. S.
Principal
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE - 410507

SL
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

21/07/22	Girls Hostels	1) All 4 floors - Rain water leaking through the window - do the wood full.	22/07/22	This work related to carpentry dept. & construction dept. So complaint forward to these dept.	SL
22/07/22	ENT OPD.	1) In our ENT OPD there are two wash basin pipes that are not working properly. to change the same.	25/07/22	work done.	SL
25/07/22	CCL.	In CCL wash room flush is not working properly. please repair as soon as possible.	27/07/22	work done.	SL
25/07/22	Laundry.	1) The pipe of wash basin in laundry damaged. please change -	27/07/22	work done.	SL
28/07/22	MMW.B.20	1) Air Net over the hall of Aluminium strips to side of corner filed.	28/07/22	These complaint forward to A (carpentry. dept.)	SL
26/07/22	MMW.B.20	There is not water coming from tap - in pte toilet	28/07/22	work done.	SL

K. S. Khandke
 PRINCIPAL
 MIMER MEDICAL COLLEGE
 TALEGAON DABHADE,
 PUNE - 410507

SL
 Maintenance Dept.
 MIMER Medical College
 Talegaon Dabhade

K. S. Khandke
 PRINCIPAL
 MIMER MEDICAL COLLEGE
 TALEGAON DABHADE,
 PUNE - 410507

SL
 Maintenance Dept.
 MIMER Medical College
 Talegaon Dabhade

Aug - 2022

1/08/22 CD & Y Dept. Breast feeding week
2nd Aug 22 to 6th Aug 22 celebrate
in sukhnd Hall Day 4 & Day 5th

4/08/22 Hall Handover to ERG Dept
cleaning work done. *SL*

3/08/22 (Girls Hostel) 1) Ground floor - faucet (spray) has
been broken needs to be changed.
(Right side Toilet No. 04)

5/08/22 work done. *SL*

2) Gs floor - wash Basin is leaking
need to be repaired.

5/08/22 work done. *SL*

3) 2nd floor reading Hall bathroom
panage checkup -

5/08/22 work done. *SL*

8/08/22 Skin & VD. 1) while shifting the glass of the
cupboard wash broken so requesting
that cupboard glass as soon as
possible.

work done. *SL*

12/08/22 M.R.D. Medical Record Department
1) MRD मध्ये, डॉक्टरांच्या सुचनेनुसार
दोस्त्यांच्या वरिष्ठांच्या वरिष्ठांच्या
शासनाच्या बाबत डॉक्टरांच्या सुचनेनुसार
संबंधित नोंद घेतली आहे. डॉक्टरांच्या सुचनेनुसार
दोस्त्यांच्या वरिष्ठांच्या वरिष्ठांच्या बाबत
दोस्त्यांच्या वरिष्ठांच्या वरिष्ठांच्या बाबत

संशोधित वसु न संशोधित विषय
दोस्त्यांच्या वरिष्ठांच्या वरिष्ठांच्या बाबत

Ksandy
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE 410507

SL
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

Ksandy
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE 410507

SL
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

12/08/22 fsw +
fow. • Taps leakage in ward water
waste.

12/08/22 1) on the occasion of Independence
Day celebration 15th Aug. 22
Choir arrangement by maint.
dept

17/08/22 Peads
Department 1) Complaint about there is a water
leakage from basin pipe. in
peads opd. A1

17/08/22 Opthal
Department Eye Donation Fortis
night celebration.
from - 25th Aug 22 to
8th sept. 2022
10 days. in OPDS wards
and cishout Hall.

26/08/22 Vkratunda
programmes Requisition for arrangement
of 25th table and one notice
board. in Auditorium

Khandys
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE,
PUNE - 410507

sls
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

12/08/22 work done sls

14/08/22 work done sls

20/08/22 work done sls

26/08/22 Hand over keys of cishout Hall to
opthel dept. sls

30/08/22 25th tables from cishout hall &
one notice Board in Auditorium done sls

Khandys
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE 410507

sls
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

जुलै - 2021

01/07/21	Wood No. 108.	Wash basin tap leakage. in patient toilet.	10/07/21	work done.	Satish Lawdhan	SL.
1/07/21	Centru Research Labs.	Kindly check the tap in Ladies Toilet.	10/07/21	work done.	Ashwan e Satish.	SL
5/07/21	Boys Hostel	i) water tank of Auto Plant & using water tank need to be cleaned.	15/07/21	work done.	M. Warkar	SL
8/07/21	0007/W/108	Patient toilet tap is leakage.	18/07/21	There No tap leakage work.	Satish	SL.
20/07/21	Casualty.	Mixer o.t. Basin is block.	24/07/21	work done.	Ashak Keshwani	SL.
20/07/21	Casualty.	Three Liquid container not working.	24/07/21	work done.	Mahadev Chetne	SL
20/07/21	Medicine Dept.	Dialysis machine & R.O. plant Not functioning will be stop.		R.O. plant ^{Need} Membrane replaced work done.	Green Enviro	SL.

PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE - 410507

PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE - 410507

PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE - 410507

PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE - 410507

21/08/21

6/08/21	Anatomy Dept	Repairing of wash Basin tap in Histo Lab. (02 wash Basin tap)	9/08/21	work done.	Phagwan Satisb	SL
9/08/21	Boys Hostel	Urine pot is checked up.	10/08/21	check-up removed	Kisan	SL
11/08/21	Peeds OPD-40	Wash Basin tap is not working properly.	11/08/21	work done.	Phagwan Satisb	SL
30/08/21	OBGY OPD-44	Wash Basin pipe is broken and need to be repaired or replanted	31/08/21	New basin pipe fit in OPD.	Satisb	SL
31/08/21	OBGY - Wood-108	1) In the labour room and side room Ceiling fungus needs to be repaired 2) Wall tiles are Broken and need to be repaired or replaced.		This complaint transfer to construction Dept.		SL
				This complaint transfer to construction Dept.		SL

PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE 410507

PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE 410507

PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE 410507

PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE 410507

Sept. 2021

2/09/21	ENT OPD	one Basin tub & Bathroom flush is not working properly.	11/09/21	work done by plumber New tap fitting.	Ajay Alok	SS
16/09/21	Blood Bank	Leakage of basin pipe in the autoclave Room.	18/09/21	New basin pipe fit to the basin autoclave Room.	Ajay Alok	SS
16/09/21	Casualty	minor o.t. side wash basin is check up.	18/09/21	work done	Ajay Alok	SS
29/09/21	operation Theater	in o.t. No. 02, and 05 water tap and soap stand needs to repair.	30/09/21	water tap repairing done. Soap stand repair.	Ajay Alok	SS

Khanduja
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE,
PUNE - 410507

SS
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

Khanduja
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE,
PUNE - 410507

SS
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

OCT. - 2021

5/10/21 Biochemistry Dept. Due to the fitting of window in U.G. laboratory, there is need of fitting of taps and repairing of wash basin.

8/10/21 Anatomy Dept. For Anatomy day functions following material in the Auditorium
 1) Chair - 10 Nos
 2) Tables - 15 Nos
 3) Big standee - 10 Nos.

11/10/21 pharmacology dept. please arrange for sanitization as per Covid protocols during the MUHS, theory exams.

11/10/21 OBGYN Ward - No-108
 1) In the ANC ward, 03 Tailor scissors are not working properly need to be sharpened.
 2) In the ANC ward, 05 episiotomy scissors are not working properly, need to be sharpened.

11/10/21 Boy Hostel Drinking water tank need to be cleaned.

8/10/21 Two tap are fit to basin and Tiles (new) fitting to wash Basin work done. (Plumber) Pujari. Bely Jadhav. SL

11/10/21 work done Maintenance staff SL

11/10/21 during the exam period. daily spray Sanitizers for staff. P. Ghinde SL

16/10/21 Sharpening done. Bely Jadhav SL

16/10/21 Sharpening done. — SL

18/10/21 work done. Maintenance staff SL

Khandys
 PRINCIPAL
 MIMER MEDICAL COLLEGE
 TALEGAON DABHADE,
 PUNE 410507

SL
 Maintenance Dept.
 MIMER Medical College
 Talegaon Dabhade

Khandys
 PRINCIPAL
 MIMER MEDICAL COLLEGE
 TALEGAON DABHADE,
 PUNE 410507

SL
 Maintenance Dept.
 MIMER Medical College
 Talegaon Dabhade

11/10/21 X-ray Toilet and checkup in X-ray Dept.

11/10/21 work done. Astak RS

13/10/21 Anaesthesia Dept. world. Anesthesia day. 16/10/21 following material required:
1) 15 poster stand.
2) Black Board. with color chalk
3) one table
4) six chair.

15/10/21 work done. Maintenance Staff RS.

19/10/21 Ward No. B-20 1) wash basin tap is not working

20/10/21 work done. New wash basin tap replaced. Satish Bhagwan RS.

24/10/21 Anatomy Dept. students exhibition program on 14th Oct. 21
1) 100 chair - (150 + 10 for faculty)
2) 10 Big standee.
3) 15 tables.
4) 2 Tree pay.

19/10/21 All material arranged by maintenance dept. RS.

20/10/21 Ward No B-106 03 Toilet & Bathroom blockage in ward.

20/10/21 work done. Kamraj RS.

21/10/21 Skin OpD Demoroom tap is not working properly.

23/10/21 Tap repaired by plumber. Satish Bhagwan RS.

Khandagale
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE,
PUNE 410507

Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

Khandagale
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE,
PUNE 410507

Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

Nov - 2021

1/11/21	SKIN OPD	phototherapy room wash Basin pipe is checkup - & demo room tap is not working.	5/11/21	New Basin pipe replaced. Tap is repairing.	Satish Bhagwan	RS.
6/11/21	R.G. GMS.	Resident GMS Hospital Right side toilet flush is not working.	8/11/21	Flush Repairing done	Satish	RS
6/11/21	I.C.U.	DO: Rooms flush is not working property.	6/11/21	New flush spare part fitting work done.	Praywan Bhude	RS
16/11/21	Ward No. B-20.	Staff Toilet wash Basin tap is not working.	23/11/21	New wash Basin tap fitting replaced.	Satish Bhagwan	RS.
17/11/21	Anaesthesia -logd dept	In operation theatre DO'S Toilet hand spray is not working.	24/11/21	New Health Faucet & Two way tap fit in Toilet	Satish Bhagwan	RS.
29/11/21	Boys Hostel	1) Ground floor - Bathrooms shower broken 2) 1st floor (Right side) 1) wash basin tap broken 2) Angle rock is jammed 3) shower has broken (Left side) 1) Inlet pipe damaged 2) Toilet flush cock jammed		All toilet taken to Renovation by construction dept.		RS.

Khandagale
PRINCIPAL
NIMER MEDICAL COLLEGE
TALGAON DABHADE
PUNE 411007

Maint. Dept.
NIMER Medical College
Talgaon Dabhade

Khandagale
PRINCIPAL
NIMER MEDICAL COLLEGE
TALGAON DABHADE
PUNE 411007

RS
Maintenance Dept.
NIMER Medical College
Talgaon Dabhade

- Boys Hostel - 5 on 2nd floor
- 1) Urinal nozzle (sprayer) are broken (02 Nos)
 - 2) Urine pipe are broken (02 Nos)
- 4) 3rd floor (Left side)
- 1) Urinal sprayer are broken (02 Nos)
 - 2) Wash basin tap is loose need to be fixed.
 - 2) Toilet flush is loose.
- (Right side)
- 1) Toilet faucet is loose
 - 2) Bathroom shower is broken
 - 3) Bathroom tap cock is broken
 - 4) cooler cock is loose
 - 5) Toilet tap is loose

All Toilets of Boys Hostel taken
to renovation by construction dept

work done (New tap fit)

SK

K. S. K. S.
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE - 410507

SK
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

K. S. K. S.
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE - 410507

SK
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

Dec. - 2021

- 17/12/21 Girls Hostel
- 1) R.O. water tanks needs to be cleaned. (02 Nos)
 - 2) Drinking water tank and other water tank needs to be cleaned (02 Nos)
 - 3) water tanks on the terrace needs to be cleaned. (3 Nos)

- 1/12/21 Girls Hostel
- 1) pipeline on the terrace is leaking
 - 2) There is Leakage under the R.O. water filter

- 06/12/21 Gynce ward No. 106 & 108
- 1) In the pnc ward No-106. four windows glass are broken and need to be replaced or repaired.

- 7/12/21 Central Re. Lab.
- 1) The urinal on the 3rd floor (Near-CRL) is leakage. please do the needful

- 11/12/21 Orthopaedics Department
- 1) Guest Lecture on Vascular Training by Dr. Atul. S. Sadhan on 13/12/21 Time 2. pm - 3 pm at Shushrut hall. cleaning

Khandogke
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE 410507

Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

- 7/12/21 work done
- 10/12/21 work done

- 3/12/21 Repairing of pipeline work done. Re plant also repaired
- Sahay
Sureshkar
Patil

- 13/12/21 04 window Glass Broken is fitting work done. (replace New Glass)
- Suresh
Gade

- 8/12/21 work done (Central checkup & lockbox)
- Rajan
Patil

- 12/12/21 work done
- Pandey
Chaitan Kumar

Khandogke
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE 410507

Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

17/12/21 Surgery Dept. 1) Regarding male surgical ward No. 205 patient toilet commode moved from their place. please repair it as soon as possible. plumber S.S.

18/12/21 physiotherapy college 1) we required a drainage pipe for the installation of constant bath unit in the musculoskeletal physiotherapy unit. please do needful. plumber S.S.

21/12/21 Gynaec ward No. 108. 1) staff toilet checkup in the ward. cook done checkup remain! S.S.

22/12/21 C.C.U 1) taps in the clinicopathology room is broken. we would kindly request repair as soon as possible. plumber S.S.

29/12/21 Blood bank 1) The water tap in the ladies toilet is leakage. repair. plumber S.S.

Khandag
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE,
PUNE-412007

S.S.
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

15/12/21 New commode fitting in the ward. plumber S.S.

inlet & out let pipeline connection given. work done. plumber S.S.

21/12/21 cook done checkup remain! S.S.

29/12/21 tap repairing. work done. plumber S.S.

30/12/21 new tap repaired work done. plumber S.S.

Khandag
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE,
PUNE-412007

S.S.
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

29/12/21 Word NO. 22 - 1) Bathroom of staff toilet is Checkup please do the work.

29/12/21 Checkup remain. Ashok walnik; SCS

29/12/21 Workshop Chestnut - 1) Duty Analysis & Referencing workshop on the sushant hall 7/Jan/22 Friday

7/01/22 sushant hall cleaning work due. Tanaji Shilanka SCS

Khandys PRINCIPAL MIMER MEDICAL COLLEGE TALEGAON DABHADE. PUNE - 410507

SCS Maintenance Dept. MIMER Medical College Talegaon Dabhade

Khandys PRINCIPAL MIMER MEDICAL COLLEGE TALEGAON DABHADE. PUNE 410507

SCS Maintenance Dept. MIMER Medical College Talegaon Dabhade

Jan - 2022.

3/01/22	Girls Hostel	1) First floor - right side (Bathroom) Glass broken.	4/01/22	New window frames replaced.	Surash Gade	RS
6/01/22	Booth Wood	1) Command in the male booth wood is broken. please do the needful.	13/01/22	Command fitting work done	Safish Sawadeka plumber	RS
10/01/22	Operation Theater	1) In front of O.P. No. 597 water tap is leakage.	10/01/22	New tap replace	plumber	RS
11/01/22	micro dept.	1) Two persons are positive so please sent the person to disinfected the dept.	11/01/22	2% sodium hypochlorite sparing in the dept.	prakash shinde	RS
15/01/22	C.C.L.	1) kindly repair the tap in C.C.L. (Biochemistry section) is leakage of tap water.	17/01/22	New tap replaced. work done.	plumber	RS
22/01/22	N2CW	1) Bathroom checkup in N2CW. please do the needful	22/01/22	work done.	Ashok Walmiki	RS

Khandyale
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE,
PUNE - 410507

RS
MIMER
Talegaon

Khandyale
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE - 410507

RS
MIMER
Talegaon

24/01/22 Girls Hostel

1) first floor & Third floor - screws of motor cover of cooler have been loosen need to be fixed.

31/01/22 Anatomy Dept

1) Replace the wash Basin tap in our department museum - wash Basin tap not working.

28/01/22 - water cooler screw fitting work done

SS

2/02/22 New Wash Basin tap replace work done

Satish Bhagwan

SS

Krandhige
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE,
PUNE - 410507

SS
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

Krandhige
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE,
PUNE - 410507

SS
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

feb - 2022

2/02/22	Oortho Ward.	1) Water supply is off in staff Toilet in male oortho ward.	3/02/22	work done.	plumber	SEL
24/02/22	Nursing collage.	1) white tube connected to the wash Basin is broken to dry. do the needful.	4/02/22	New Basin pipe replaced work done.	plumber	SEL
7/02/22	Physiology Dept.	1) one tap of physiology dept. 2) Ladies Toilet Leakage water overflow in the Toilets.	8/02/22	New tap replace. Repairing of flush tank work done.	plumber	SEL
8/02/22	P.G. Hostel	D.R.Q. (Hospital) P.G. girls Hostel first No. right side Toilet flush is not working.	10/02/22	work done	plumber	SEL
9/02/22	PNC Ward-No. 106	4 Bathrooms checkup please do the needful.	9/02/22	work done.	Ashwini	SEL
19/02/22	PSY Ward.	2 bed Bathrooms window glass are broken replace urgently.	21/02/22	All window glass fitting work done.	Sunandha Gowde	SEL

Ksundhya
PRINCIPAL
MIMER MEDICAL COLLEGE
TALGAON DABHADE,
PUNE 410507

SEL
Maintenance Dept.
MIMER Medical College
Talgaon Dabhade

Ksundhya
PRINCIPAL
MIMER MEDICAL COLLEGE
TALGAON DABHADE,
PUNE 410507

SEL
Maintenance Dept.
MIMER Medical College
Talgaon Dabhade

18/02/22	College	please give tables & Notice Board etc for celebrating of shiv Sanyas;	18/02/22	work done		SL
25/02/22	Ortho ward	1) flush in the patient toilet is not working 2) There is no toilet washing pipe in the same toilet 3) There is no doctor tap in wash Basin please do the needful	25/02/22	Taking to the Regulation in month. so not work done.		
			25/02/22	work done.	plumber	SL
			25/02/22	work done.	plumber	SL
25/02/22	PSY Dept	workshop on sleep Awareness week, please request allowed to use Sushrut Hall on 16th March 2022, from 2.30 pm to 4 PM.	16/03/22	Shiv Sushrut Hall handover to psy dept. on 2.00 pm.	Tanaji	SL
28/02/22	Surgery Dept	14th Annual Hands on Trainee Endoscopy workshop.	2/03/22	Sushrut Hall handover to Surgery Dept. Hall clearly done	Tanaji	SL

Khandys
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE,
PUNE - 410507

SL
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

Khandys
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE,
PUNE - 410507

SL
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

March - 2022

1/03/22	D.S. Darganji Sushrut Hall	A meeting for introduction and reorientation to the various existing health related schemes and services at Dr. B.S.R Hospital will be conducted on 02nd march in Sushrut Hall.	2/03/22	provide sru sushrut Hall ans 2:00 pm. to 4:00 pm.	Tanaji S.S.
1/03/22	sushrut Hall Physiotherapy College.	1) Kinesiocon - 2022 E-conference on 7th & 8th April 2022. from 8:00 am to 4:30 pm. sushrut Hall in opd completed.	6/04/22	hand over sushrut Hall to physiotherapy Balu Sode S.S. College.	
3/03/22	skin & VD OPD.	D our water tap is repairable, please do the needful as early as possible.	9/03/22	Two tap of wash Basin repaired. work done. Hand cleaning done.	plumber S.S.
8/03/22	Boys Hostel	R.O. water tank (white colour) needs to cleaned.		work done	S.S.
8/03/22	FCU.	pipe of wash Basin is Broken in FCU please repair it urgently.	9/03/22	New wash Basin pipe checked	plumber S.S.

Krandhga

PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE,
PUNE 410507

Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

Krandhga

PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE,
PUNE 410507

Maintenance Dept
MIMER Medical College
Talegaon Dabhade

9/03/22 ENT OPD The water tap not functioning properly in ENT OPD. Kindly repair it.

9/03/22 Physiotherapy College 1) The out let pipe of water cooler needs to be changed.
2) In girls toilet Bathroom out let pipe needs to be changed.

11/03/22 Girls Hostel Ground floor right side lavs Nos. Toilet flush is not working.

12/03/22 C.C.L Kindly repair the tap in C.C.L (Microbiology section) water tap leakage.

16/03/22 Opthal hood. Male & female opthal hood washer colour checkup

Khandya
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE
PUNE-410507

SS
Maintenance Dept.
MIMER Medical College
Talegaon DabhaDe

10/03/22 work done. plumber SS

14/03/22 1) New pipe fitting. work done plumber SS

2) New pipe fitting work done plumber SS

14/03/22 work done flush repair plumber SS

15/03/22 New tap fitting work done by plumber plumber SS

16/03/22 work done checkup removal ASBOK SS

Khandya
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE.
PUNE-410507

SS
Maintenance Dept.
MIMER Medical College
Talegaon DabhaDe

17/03/22	Staff Gr. No. 4 Acwing.	Repair of Toilet flush pipe in staff Gr. No. 04.	17/03/22	work done	Cetish	SL
19/03/22	RHTC Sudumbare	1) first floore Room No. 4 female Interns Gr. through the water pipe line is there not a single drop of water available at Basin, Bathrooms, Latrine, 2) Gr. floore Room No. 06 (lab) 1) Basin Has been splintered need to be replaced 2) Cock has been damaged needs to be replaced.		work done	plumber Satish & Raghuvaran	SL
21/03/22	PNC Wood	2 Bathroom blocked & cooler cooler blocked.	21/03/22	work done Checkup removed.	Ashak A. Jay.	SL
21/03/22	ENT Department	please provide suchrat Hall for Lecture from 11.00 am to 2.00 pm today & tomorrow, 21/03/22 & 22/03/22	22/03/22	Water Checkup repair work done	Ashak	SL
21/03/22			21/03/22	Handover Key of suchrat Hall after cleaning work done	Fungy!	SL

Khandys
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE,
PUNE 410507

SL
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

Khandys
PRINCIPAL
MIMER MEDICAL COLLEGE
TALEGAON DABHADE,
PUNE 410507

SL
Maintenance Dept.
MIMER Medical College
Talegaon Dabhade

DATE	No of Inwarded LED Tube light	From	No of Outwards LED Tubelight
16/05/22	01 LED Box inwarded 200 qty	central Store	(20 qty)
30/05/22	01 LED Box inwarded	central Store	20 qty
01/06/22	10 Wanda LED inwarded	— 11 —	03 qty
16/05/22	-	-	01 LED Tube fixed
17/05/22	-	-	01 LED Tube fixed
18/05/22	-	-	01 LED Tube fixed
20/05/22	-	-	02 LED Tube fixed
21/05/22	-	-	03 LED Tube fixed
23/05/22	-	-	01 LED Tube fixed 01 LED Tube fixed
26/05/22	-	-	01 LED Tube fixed
28/05/22	-	-	04 LED Tube fixed
30/05/22	-	-	02 LED Tube fixed

IDS	Department	Name of Singer
INSIPART 011	REMARKS: No old	Singer
INSIPART 011	REMARKS: No old	sr. 2010
Community Medicine dept	old chole need to Replace by New one	Kiran Ranjit
C.C.L.	————— —————	Avinash
PTCU	old chole	Tukaram
Central Library	————— —————	Nandkisan Kiran
C.C.L.	old chole need to Replace by New one	Nandkisan Kiran
Surgery dept Girls Hostel Room No-206	————— —————	Kiran
Girls Hostel 205	old chole	Avinash
Urban center	old chole need to Replace by New one	Wairaje Kiran Ranjit
Surgery dept	————— —————	Bairam Avinash Ranjit

DATE	No of Inwarded LED Tubelight	From	No of Outward LED Tubelight
31/05/22	-	-	11 LED Tube fixed
03/06/22	-	-	07 LED Tube fixed
01/06/22	-	-	03 LED 10watt fixed (10watts)
04/06/22	-	-	02 LED Tube fixed 01 LED Tube fixed
05/06/22	-	-	02 LED Tube fixed
* 02 LED Box completed	* 40 Nos completed		
03 LED 10watt completed	03 Nos (10watt)		

By D

Boy

Amc

Pdth

Soc

M

qu

Pa

Cus

Peo

si

Pre

Date	Department	Remarks	Name & Sing
10/10/2021	Boys Hostel = 02	old Chalk need to	Manish
10/10/2021	Anatomy department = 07 Pathology department = 02	Replace by New one	Anshom Bhairava
10/10/2021	Social worker = 03		Avinash
10/10/2021	Microbiology = 02 Surgeon Ward = 02		Pravhyanshi Bhairava
10/10/2021	Pathology dept	New Required	Wairnge
10/10/2021	Cuscuth = 01 Pediatric = 01	old Chalk	Manish Munde
10/10/2021	SIR Fire Pump	old Chalk New Required	Manish Pravin

DATE	No of Installed LED Tubelight	From	No of Outlets LED Tubelight
06/06/22	1 LED Box installed	Central Store	20 Qty
13/06/22	02 LED Box installed	Central Store	40 Qty
06/06/22	-	-	01 LED Fixed
09/06/22	-	-	02 LED Fixed
10/06/22	-	-	01 LED Fixed
12/06/22	-	-	11 LED Tube Fixed
18/06/22	-	-	01 LED New Fixed
15/06/22	-	-	04 LED Fixed
17/06/22	-	-	11 LED Fixed
19/06/22	-	-	10 LED Fixed
02/06/22	-	-	03 LED Fixed
03/06/22	-	-	04 LED Fixed

Work Order / Department / Location	Remark / To / From	Name
1st Hostel (108)	108	Sing
GYN Dept	old chalk	Tukaram
Girls Hostel Staff. Q.	old chalk need to Replace by New one	Avinash Pravin Hankishan
ENT OPD	old chalk need to Replace by New one	Wairaje Baliram
Boys Hostel		Kiran
skin OPD	old Chalk	Munde Wairaje
NZU & PTGU	old Chalk to need to Replace by New one	Badi Ram Nihh Avinash
GYN-108		
New Camera Room	New Required	Nihh Avinash
Pathology Dept	old chalk need to Replace by New one	Nihh Avinash

DATE	No of Inwarded LED Tube light	From	No of Outwards LED Tube light
24/06/22	-	-	04 LED Fixed
25/06/22	-	-	07 LED Fixed
26/06/22	-	-	01 LED Fixed
* 03 LED Box Completed *			60 NOS Completed

Date	Department	Remarks	Name
12	Microbiology Dept	old chalk need to Replace by new one	Nikhil
1	1) Girls Hostel = 02 2) Pediatric Ward = 01 3) GYN Ward = 01		Nikhil
1	Microbiology Dept = 03 College 3rd Floor Passage = 04 Ophthal OPD = 01		Ranjit Bali Ram

DATE	No of Installed LED Tube light	From	No of Outstanding LED Tubelight
27/06/22	01 LED Box Installed	Central Store	20 QTY
28/06/22	02 LED Box Installed	Central Store	40 QTY
27/06/22	-	-	20 LED New fixed
28/06/22	-	-	10 LED New fixed
30/06/22	-	-	03 LED New fixed
01/07/22	-	-	04 LED New fixed
02/07/22	-	-	07 LED New fixed
04/07/22	-	-	01 LED New fixed
05/07/22	-	-	06 LED New fixed
*	* 03 LED Box completed	*	60 NOS LED overalls completed

Department	Remarks	Name
Physiotherapy College	Old Chalk need to Replace by New one	Baliram Kiran
OT = 02, 6 Phdruad OPD = 01		
Medicen OPD = 04 Sursagar = 03 OPD	—————/1—————	Avinash Baliram Kiran
X Ray = 04		
Anatomy dept	old Chalk need to Replace by New one	Nandkishan Ranjit
PNC-106 = 07	—————#—————	Nitin Avinash Wairam Nitin
Ward 108 = 01		
River = 08	old Chalk need to	
Prayer Room = 03	Replaces by New one	Avinash Ranjit
Girls Hostel = 3rd Floor Room = 05	old tube changed	Tukaram

DATE	No of Installed LED Tubelight	Power	No of Outwards LED Tubelight	Det
01/07/22	Sensor LED in model 20 Watt	central store	30 NOS	
01/07/22	-	-	11 LED Sensor fixed	Col Pass
01/07/22	-	-	14 LED Sensor fixed	all

Department	Remarks	Name
Collage B. Passage	Remarks in an	ging
APD Bidding Passage	Remarks	Buliram Kiran

Collage B. Passage to Library
Passage

Remarks Required

Nikhil
Avinash

APD Bidding Passage

Remarks Required

Buliram
Kiran

DATE	No of Inwoded LED Tubelight	from	No of Outwoded LED Tubelight	
06/07/22	01 LED Box inwoded 00 Watt	Central Store	00 Qty	
15/07/22	01 LED Box inwoded 00 Watt	Central Store	00 Qty	
06/07/22	-	-	01 LED New Fixed	
07/07/22	-	-	06 New LED Fixed	AM 47
08/07/22	-	-	02 New LED Fixed	GI
14/07/22	-	-	04 New LED Fixed	Wc
19/07/22	-	-	02 New LED Fixed	SI
19/07/22	-	-	02 New LED Fixed	0
22/07/22	-	-	01 New LED Fixed	
22/07/22	-	-	02 New LED Fixed	
22/07/22	-	-	02 New LED Tube fixed	

Department	Remarks	Name & Sig
GYN dept	old Chok	Avinash
Anatomy dept = 06 GYN - 108 = 01	New Requirad 4 old Chok	Nikh Avinash
Girls Hostel = 02	old Chok need to Replace by New one	Ranjit Balipam Pradip
Ward - 205 = 4 Skin OPD = 2	" " old Chok	Nandkishor Varninge Nandkishor Pradip
OT Passage = 2	old Chok need to Replace by New one	Avinash Varninge Avinash
Blood Bank = 1 Arban center = 2	old Chok	Nandkishor Dutray Ranjit
UV Machine Blood Bank	old Chok need to Replace by New one	Nandkishor Varninge

DATE	No of Inwarded LED Tubelight	From	No of outwards LED Tubelight
26/07/22	-	-	02 New LED fixed
28/07/22	-	-	04 New LED fixed
29/07/22	-	-	04 New LED fixed
03/08/22	-	-	01 New LED fixed
04/08/22	-	-	06 New LED fixed
08/08/22	-	-	01 New LED fixed
* 02 LED Box completed			40 Nos Tube LED down completed
28/07/22	10 watt LED Tube	2 feet = 07 Nos Central Store	
28/07/22	-	-	07 New LED fixed

Department	Remark	Name & Stry
Boys Hostel = 01 Admin office = 01	old chalk	Maringe Avinash Baliram
Hospital B. = 4	old Chalk need to Replace by New one	Avinash Kiran
Biochemist = 01 Girls Hostel = 03		Maringe Avinash Pravin
Boys Hostel River	old Chalk	Baliram Nandkishor Baliram
OPD.B. Physiotherapy = 6	old Chalk need to Replace by New one	Nandkishor Baliram
Laboratory = 1	old chalk	Nandkishor Baliram
3rd Floor Dr. Gunten Toilet & Bath Room	old Chalk need to Replace by New one	Nitin, Brajip Kiran, Avinash Pravin

DATE	No of Inwarded LED Tubelight	From	No of Outwards LED Tubelight
10/08/22	01 LED Box Inwarded 20 watt	Central Store	20 Off
23/08/22	10 watt LED Inwarded	Central Store	05 New
10/08/22	-	-	02 New LED fixed
12/08/22	-	-	01 New LED fixed
15/08/22	-	-	04 New LED fixed
17/08/22	-	-	03 New LED fixed
18/08/22	-	-	02 New LED fixed
19/08/22	-	-	02 New LED fixed
25/08/22	-	-	01 New LED fixed (10 watt)

Department	Remarks	Name of
Girls Hostel = 02	Old Chalk	Avinash, Bali Ram
Boys Hostel = 01	old Chalk need do Replacce by New one	Kiran
College B. and floor Passage	//	Kiran, Pradip Ranjit, Brajin
Urban Center	old Chalk	Pradip, Kiran Nidhi, Avinash
C.C.L. = 01 Medice opd = 01	old Chalk need do Replacce by New one	Nidhi, Kiran Ranjit, Nandya Waringe
Staff R.O. A-3 = 01 New Required cabin = 01	old Chalk New Required	Avinash Pradip
Dr Gunder 3rd floor	old Chalk	Nayide Pradip Avinash Ranjit

DATE	No of Inwarded LED Tubelight	from	No of Outward LED Tubelight
26/08/22	-	-	02 New LED Fixed
30/08/22	-	-	02 New LED Fixed
02/09/22	-	-	01 New LED Fixed
05/08/22	-	-	01 New LED Fixed
	*01 LED Box completed		00 Qty completed

Department	Remark	Name of Shop
GYN-108 Passage = 01 Nieu = 01	old chok need to Replace by new one	Nandkishor Avinash
ABGy Ward = 01 oditorium Passage = 01	//	Sind, kiran Waringe Avinash
Sgh MOD = 01	old chok	Ransit Tilkanam
C.C.L. Lab =	old chok need to Replace by New one	Avinash Nikhin

DATE	No of Inwarded LED Tubelight	From	No of Outward LED Tubelight
01/09/22	01 LED Box Inwarded (20 Qty Inwarded)	Central Store	20 Qty
05/09/22	-	-	02 New LED Fixed 01 New LED Fixed
07/09/22	-	-	02 New LED Fixed
08/09/22	-	-	01 New LED Fixed
09/09/22	-	-	03 New LED Fixed
10/09/22	-	-	01 New LED Fixed
15/09/22	-	-	02 New LED Fixed
20/09/22	-	-	02 New LED Fixed
21/09/22	-	-	03 New LED Fixed

7F
PK
P.S
P.A
(Ra

S
↑

6
6

Department	Remark	Name of Sins
7 Floor Hospital Passage = 02 Psy Ward = 01	old chalk	Nandkishor Rami's Baliram
Pst Ward = 01 P.G. Student (5 floor) = 01 (Room No-3)	old chalk need to Replace by New one	Baliram Avinash
OT	old chalk	Nikhin Waringe
GYN-108 = 02 Girls Hostel = 01	old chalk need to Replace by New one	Avinash, Kiran Waringe, Munde
Psy Ward = 01	old chalk	Borvin Baliram
Surgery dept = 01 MRD dept = 01	//	Rami's, Baliram Nandkishor
Girls Hostel = 02	old chalk need to Replace by New one	Waringe Munde
Anatomy dept = 03 (Histo Lab = 01, Demo Room = 01 - Speci FP Room = 01)	//	Avinash Baliram

DATE	No of Inwarded LED Tubelight	From	No of Outward LED Tubelight	
28/09/22	-	-	02 New LED Fixed	01 X
04/10/22	-	-	01 New LED Fixed	Boy (Proc)
* 01/10/22	* 01 Box LED Completed	* 20 Qty Completed		
06/10/22	01 LED Box 20 Watt Inwarded (20 Qty Inwarded)	Central Store		
31/10/22	01 LED Box 90 Watt Inwarded (20 Qty)	Central Store		
07/10/22	-	-	01 LED Fixed	01
10/10/22	-	-	01 New LED Fixed	Pa
12/10/22	-	-	02 New LED Fixed	m
13/10/22	-	-	03 New LED Fixed	

Department	Remarks	Name of
OPD = 01 XRay = 01	old chalk	Nithin Pradip Avinash
Boy Hostel = 01 (Room No - 226)	old chalk need to Replace by New one	Avinash Shinde
College B. Library = 01	old Chalk	Avinash
Pathology = 01	11	Tukarun Avinash
Microbiology Hood = 01 Ward - 102 = 01	old Chalk need to Replace by New one	Ranjid Avinash, Ranjir Waringe
Girls Hostel = 01 OP. Psy OPD = 01 Girls Hostel = 01	9	Nithin Baliram

DATE	No of Inwarded LED Tubelight	From	No of Outward LED Tubelight	
15/10/22	-	-	06 New LED fixed	MR PH
19/10/22	-	-	03 New LED fixed	Wd
20/10/22	-	-	03 New LED fixed	Si
27/10/22	-	-	02 New LED fixed	OP
28/10/22	-	-	03 New LED fixed	OP
02/11/22	-	-	05 New LED fixed	Wd
04/11/22	-	-	02 New LED fixed	OP
05/11/22	-	-	01 New LED fixed	S
06/11/22	-	-	01 New LED fixed	C
09/11/22	-	-	01 New LED fixed	P
11/11/22	-	-	04 New LED fixed	Pa

Department	Remarks	Name of Staff
MRD = 03 Physiotherapy College = 03 Ward-102 = 03	old Chole need to Replace by new one // //	Nandkishor Nishin Wanjare Ramjit
Girls Hostel - 3 Optical OPD - 1 OT - 1 OPD Tibet = 03	old Chole // // old Chole need to Replace by New one	Kiran Avinash Kiran Avinash Kiran Shinde
Ward-18 = 02, Surgery dept = 01 Ward 804 = 01 Biomedical dept = 01	// // old Chole	Ranjit Baliraj Nandkishor Avinash
OT Passage = 01 Boys Hostel = 01 Saldumbore Hospital = 01	// // old Chole	Avinash Kiran, Ganesh Wanjare
Pediatric OPD = 01	old Chole need to Replace by new one	Nandkishor Avinash
Pathology dept = 02 TB OPD = 01 Male ortho Ward = 01	//	Wanjare Avinash, Ranjit Kiran, Pravin

DATE	No of Inwarded LED Tubelights	From	No of Outward LED Tubelights	①
13/11/22	-	-	02 New LED fixed	M
*	02 LED Tube Box completed	*	40 Nos completed	*
= 09/11/22	01 LED Box inward (20124)	Central Store		

Department	Remark	Assesssing
Mule shoe wearer	old shoe need to Replace by New one	Warning Nondissem

Sr no.	Date	Department	Equipment name
1	1/6/2021	Physiotherapy	BP app
2	3/6/2021	OT	endoscopy
3	3/6/2021	NICU	SPO2 probe
4	7/6/2021	Ophthal OPD	Vision drum
5	7/6/2021	Casualty	Monitor
6	10/6/2021	Microbiology	Incubator
7	11/6/2021	OT	Warmer
8	14-7-21	CSSD	Sealing machine
9	15-6-21	NICU	Syringe Pump
10	18-6-21	NICU	Syringe Pump
11	23-6-21	Microbiology	Autoclave
12	23-6-21	MMW	BP app
13	23-6-21	FMW	BP app
14	23-6-21	Physiotherapy	Ultrasound quality
15	24-6-21	OT	ETCO2 CABLE
16	25-6-21	PNC	Baby wt machine
17	29-6-21	Physiology	ECG Cmachine
18	30-6-21	OT	Bi-polaer cable

Sr no.	Date	Department	Equipment name
1	2/7/2021	Peads OPD	BP APP
2	2/7/2021	PNC	BP APP
3	2/7/2021	obs 108	NST Machine
4	2/7/2021	Physiotherapy	IFT
5	5/7/2021	Covidwd	Monitor
6	7/7/2021	ECG OPD	ECG Machine
7	10/7/2021	Ophthal wd	BP APP
8	12/7/2021	MSW 205	BP APP
9	13-7-21	Micro	Autoclave
10	17-7-21	OT CSSD	ETO
11	20-7-21	OT	Tourniquet
12	21-7-21	ENT OPD	Audiometer
13	24-7-21	Casualty	Monitor
14	26-7-21	NICU	Pulse oxymeter
15	23-7-21	OBGY OPD	NST Machine
16	24-7-21	MMW	BP APP
17	27-7-21	PNC	BP APP
18	28-7-21	NICU	Warmer
19	29-7-21	OT	Drill machine
20	29-7-21	ENT OT	Debrider

Sr no.	Date	Department	Equipment name
--------	------	------------	----------------

1	2/8/2021	NICU	Monitor
2	2/8/2021	NICU	warmer
3	2/8/2021	OBS 108	BP App
4	2/8/2021	ICU	Nebulizer
5	2/8/2021	ENT OT	Drill
6	4/8/2021	Casulty	BP App
7	7/8/2021	PICU	Syringe pump
8	9/8/2021	Physiotherapy	CPM M/C
9	10/8/2021	Casulty	Monitor
10	11/8/2021	FMW	Nebulizer
11	10/8/2021	MMW	BP App
12	13-8-21	NICU	warmer
13	13-8-21	OBS 108	Monitor
14	17-8-21	NICU	Monitor
15	17-8-21	OBS 108	NST
16	26-8-21	OBS 108	Monitor
17	27-8-21	NICU	warmer
18	30-8-21	ICU	BP App
19	30-8-21	Pead ward	BP App

Sr no.	Date	Department	Equipment name
1	8/9/2021	OT	Central Nitrus
2	8/9/2021	Radiology	X-ray
3	11/9/2021	NICU	Central suction
4	14-9-21	MedicineOPD	BP App
5	8/9/2021	OT	Stryker trolley
6	18-9-21	Ophthal OPD	Vision drum
7	18-9-21	Ophthal OPD	Slit lamp
8	20-9-21	OBGY OPD	NST M/C
9	16-9-21	CASULATY	RMS Multipara monitor
10	20-9-21	OT	OT Table
11	22-9-21	NICU	Neonatl laproscope blade
12	22-9-21	PNC Ward-106	pulseoximeter
13	24-9-21	ophthal ward	BP Apparatus
14	28-9-21	casulty	laryngscope blade
15	30-9-21	PNC Ward-106	Nebuliser machine
16			
17			
18			
19			
20			

Sr no.	Date	Department	Equipment name
1	4/10/2021	PICU	Suction machine
2	4/10/2021	OBS-108	BABY warmer
3	4/10/2021	ICU	BP AAPARATUS
4	5/10/2021	NICU	SPO2
5	7/10/2021	FMW-22	BP AAPARATUS
6	7/10/2021	Microbiology	electrical balance
7	7/10/2021	ICU	Vision laryngoscope
8	7/10/2021	OBGY OT	Light source
9	8/10/2021	skin&minor OT	MPA MACHINE
10	8/10/2021	NICU	Syringe pump
11	11/10/2021	OT	Cystoscopic sheath
12	12/10/2021	PICU	Needal cutter
13	12/10/2021	NICU	Central suction bottle,o2 bottle
14	12/10/2021	NICU	Bellasta ventilation humicition fisher&payer
15	12/10/2021	NICU	Triviron ventillation humidifier CH580
16	12/10/2021	CCL(BIOCHEMISTRY)	CENTRIFUGE
17	18-10-21	OBS 108	NST MACHINE
18	18-10-21	Psy ward	weight machine
19	20-10-21	OBS 108	BABY warmer
20	21-10-21	FMW-18	CENTRAL O2 PORT Maltipara monitor,spo2 probe,laryngoscope,
21	21-10-21	ICU	ballon of BP
22	22-10-21	OBS-108	O2 METER
23	23-10-21	OBS WARD	BP AAPARATUS
24	24-10-21	OBS WARD	NST MACHINE
25	26-10-21	OT ORTHO	MANMAN DRILL MACHINE
26	26-10-21	PNC WARD 106	BP AAPARATUS
27	29-10-21	casulty	r-m-s multipara monitor

Sr no.	Date	Department	Equipment name
1	1/11/2021	m.s.w	bp aap.
2	3/11/2021	ot cssd	ETO machine compresor
3	3/11/2021	ot cssd	autoclave machine
4	5/11/2021	microbiology	weighing machine
5	8/11/2021	pnc ward	bp aap.
6	10/11/2021	F.S.W	bp aap.
7	12/11/2021	ICU	2 SYRINGE pump
8	16-11-21	M.O,W	bp aap.
9	16-11-21	F.S.W	Needle cuter
10	17-11-21	NICU	Central O2
11	19-11-21	ICU	bp aap.
12	19-11-21	ICU	Defib (old)

13	19-11-21	casualty	bp aap.
14	22-11-21	NICU	multipara monitor
15	22-11-21	NICU	Baby warmer
16	22-11-21	NICU	O2& Suction
17	22-11-21	OBS 108	bp aap.
18	23-11-21	M.O.W	bp aap.
19	24-11-21	Dailysis	weight machine
20	24-11-21	OBS 108	Pediatric laryngoscope
21	26-11-21	PICU	Central O2
22	26-11-21	Anatomy	bone cutting machine
23	27-11-21	M.M,W	BP app
24	27-11-21	ICU	multipara monitor
25	27-11-21	medicine opd	bp app
26	30-11-21	NICU	SPO2 PROBE
27	30-11-21	microbiology	autoclave machine

Sr no.	Date	Department	Equipment name
1	6/12/2021	ophthology opd	vision drump remote
2	6/12/2021	obs 108	baby warmer
3	7/12/2021	NICU	Not working
4	8/12/2021	dental opd	confident dental chair
5	8/21/2021	ICU	syringe pump
6	15-12-21	dialysis	dialusis machine
7	16-12-21	icu	bp app
8	17-12-21	microbiology rtPCR lab	de freeze
9	18-12-21	icu	nebuliser omron
10	20-12-21	ot	ot no 4 light glass
11	23-12-21	casualty	multipara monitor
12	23-12-21	nicu	central suction
13	24-12-21	fmw +paed ward	needle cutter
14	24-12-21	fmw ward	bp app
15	27-12-21	icu	central suction
16	27-12-21	physiotherapy	infra red 2
17	28-12-21	radiology	wipro ge dx -525
18	29-12-21	casualty	weight machine
19	29-12-21	dental opd	dental chair compressor

Brakedown	Action taken (Inhouse/Outsource)	Remark
ballon torn	Inhouse	
Softwareproblem	Outsource	
sensor faulty	Inhouse	
remot faulty	Inhouse	
Display not working	Inhouse	
Not working	Inhouse	
notgetting heat	Inhouse	
Not getting heat	Inhouse	
Not working	Inhouse	
Not working	Inhouse	
coil burnt	Inhouse	
Lekage	Inhouse	
Not working	Inhouse	
connection broken	Inhouse	
Not working	Outsource	
Cable broken	Inhouse	
Battery faulty	new given	
Not working	Inhouse	

Jul-21

Brakedown	Action taken (Inhouse/Outsource)	Remark
Not working	Inhouse	
Not working	Inhouse	
button broken	Inhouse	
Broken	Inhouse	
Display problem	Not repairable	
Button not working	Inhouse	
Not working	Inhouse	
Murcury tube broken	Inhouse	
Coil burnt	Outsource	
Motar burnt	Outsource	
Display problem	Inhouse	
Not working	Inhouse	
wire burnt	Inhouse	
Display problem	Outsource	
Not working	Outsource	
LEAK	Inhouse	
Not working	Inhouse	
not getting heat	Inhouse	
Not working	Outsource	
Not working	Outsource	

Aug-21

Brakedown	Action taken (Inhouse/Outsource)	Remark
------------------	---	---------------

Not working	Inhouse
Not working	Outsource
Not working	Inhouse
Not working	Inhouse
hand pice problem	Outsource
Leakage	Inhouse
Not working	not repairable
Plug shocket broken	Outsource
Diplay not showing	Outsource
Not working	not repairable
Mercury tube broken	not repairable
not getting heat	Inhouse
Diplay not showing	Inhouse
charging problem	Inhouse
PPE Roll button faulty	Inhouse
Diplay not showing	not repaired
Not working	Outsource
Not working	Inhouse
Not working	Inhouse

Sep-21

Brakedown	Action taken (Inhouse/Outsource)	Remark
Leakage	Outsource	
Wheels broken	Outsource	
blockage	not repaired	
Bollon damage	Inhouse	
light source not working	Outsource	
Remot working		
Foot padal not working	Outsource	
Door problem	Outsource	
screen not waving		
not working properly	Outsource	
not working proply		
pulseoxmeter broken	spring broken which is not repairable	
not working properly	replace rubberbag,button,value	Inhouse
scope & blade not working	Inhouse	
machine nob is broken	not repairble	

Oct-21

Brakedown	Action taken (Inhouse/Outsource)	Remark
not working properly very low pressure	repair	
not working properly	need to cyramic coil change	
Button not working	repair	
Not working properly	repair	
not working proprally		
not working	outsource	
NOT working proprally		
not working	not repair	
cystoscopic sheath broken ¬ working		
not working	not repairable	
o2 bottle not working		
humidifier not working	not repairable	
humidifier not working	not repairable	
Not working ,parts are broken	not repairable	
Not working	not repair	
Cyramic coil not working	not repair	
NOT working	not repairable	
ALL equipment not working	only ballon of bp working,	
O2 METER not working	not repairable	
NOT working	repair	
NOT WORKING	not repair	
Drill machine not working ,drill damage	outsource	
Mercury is not going up after pumping also on off problem ,screen blinked	not repair	

Nov-21

Brakedown	Action taken (Inhouse/Outsource)	Remark
hinjus breaking	change the hinjus	repair
not working	inform to company person	outsource
need to servicing	company person came &check need to replace coil	outsource
not working	check it	repair
not working	baloon change	repair
baloon not working	baloon change	repair
not working	check it	repair
baloon not working	baloon change	repair
syringe not cutting	burner jaming check it	repair
O2 leakage problm	check out it need to replace o-ring	
baloon not working	change the baloon	repair
not working		

baloon not working	baloon change	repair
bp reading not showing	check it	repair
reading not showing	check the button	repair
O2 Leakage problm& Suction point not working		
BP App broken	glass tube change	repair
not working	baloon change	repair
not working	giving to new	
not working		
Leakage problm		
it is not functioning	company person came &check it waiting the approval	
baloon not working	baloon change	repair
not working		
not working	replace rubber bag	repair
not working	Check it	repair
air leak	check it	repair

Dec-21

Brakedown	Action taken (Inhouse/Outsource)	Remark
not working		repair
not working	coil replaced ,power cord replaced	repair
display problm	waiting for approval	
not working	company person came & check	repair
Main power cord switch burn		
level detector faulty	need to replace sensor	outsource
broken		repair
not working		
not working	open that machine and check out	repair
light glass is broken		outsource
csreen not working display not shown		
is not working		
not working		repair
not working		repair
not working		
plug broken		
machine not working	engee. Camw and repair	outsource
not working		
machinecreating to much noise		

sr	date	dept	equipment name
1	3/1/2022	obgy opd	2 bp app
2	4/1/2022	m&f opthal ward	bp app
3	4/1/2022	pnc ward	bp app
4	8/1/2022	microbiology (rtPCR)	deep freeze
5	10/1/2022	labour room	bp app
6	11/1/2022	ent ot	bean spliter opthal
7	13-1-22	fsw	multipara monitor
8	14-1-22	obgy opd	bp app
9	15-1-22	casualty b-1	phobius rms monitor
10	17-1-22	obs 108	o2 mater
11	18-1-22	obs 108	adult weight machine
12	17-1-22	ot	syringe pump
13	19-1-22	fmw	monitor global medicare
14	19-1-22	ot	central suction ,spo2 probe,ortho.manman foot switch
15	20-1-22	nicu	central line o2 leakage
16	20-1-22	fsw	o2 flow meter
17	20-1-22	nicu	laryngscope blade
18	20-1-22	nicu	warmer meditrion
19	5/1/2022	picu	emco monitor
20	27-1-22	blood bank	hemocare machine hb
21	27-1-22	nicu	humidifier bottle ,flow meter
22	28-1-22	microbiology (rtPCR)	remi freezer 80c
23	29-1-22	pnc 106	needle cutter
24	29-1-22	dental opd	dental chair,suction machine ,scalar unit
25	29-1-22	ot	laryngscope blade pead,cautry cables 2
26	31-1-22	ecg	ecg machine print not shown,lead not working
27	31-1-22	ot	orange colour cable ,syringe pump

sr no	date	dept.	equipment
1	1/2/2022	nicu	ge warmer
2	2/2/2022	pnc ward	bp app, pead weight machine
3	2/2/2022	opthal ward	perimeter
4	1/2/2022	nicu	ventilator bellavista
5	3/2/2022	opthal ot	autoclave ot m.shah
6	3/2/2022	blood bank	bp app,
7	8/2/2022	opthal opd	vision drum
8	8/2/2022	icu	2 monitor
9	8/2/2022	ortho ot	manman cable,hp, remaing hp
10	9/2/2022	casulty	bp app

11	9/2/2022	fsw	bp app
12	9/2/2022	nicu	trivitron ventiii
13	10/2/2022	physiotherapy	bp app, ift mach.electrodes
14	14-2-22	pnc ward	digital spo2
15	14-2-22	obs 108	bp app
16	18-2-22	mow	bp app
17	18-2-22	icu	2 bp app
18		icu	resmed ventii
19	25-2-22	nicu	neonetal spo2 trivitron
20	25-2-22	microbiology	bactec machine
21	26-2-22	dialysis	1 bp monotor 2 nipro dialysis
	28-2-22		

sr no	Date	dept	equipment
1	1/3/2022	icu	ventii drager
2	3/3/2022	icu	old spo2 m/C covidien total 3
3	4/3/2022	nicu	needle cutter
4	8/3/2022	pnc ward	spo2 probe digital
5	9/3/2022	obgy opd	weight m/C
6	8/3/2022	casulty B-1	Resmed ventii
7	9/3/2022	opthal opd	slit lamp zais
8	11/3/2022	dialysis	dialysis m/C
9	14-3-22	opthal ot	carl zeiss microscope
10	14-3-22	msw	bp app
11	15-3-22	pnc ward	multipara monitor
12	16-3-22	ot	ortho ot table gynac hysteroscope
13	17-3-22	nicu	ventilator
14	18-3-22	casualty	bp app
15	19-3-22	casualty	2 phobius monitor
16	19-3-22	nicu	covidein spo2
17	21-3-22	pnc ward	bp app
18	21-3-22	nicu	neonetal spo2 probe
19	22-3-22	mow 205	bp app
20	24-3-22	obgy opd	2 led lamp
21	24-3-22	fmw	needle cutter
22	26-3-22	nicu	radient warmer ,multipara monitor

SR	DATE	DEPT	EQUIPMENT
1	4/4/2022	Peadatric ward	bp app
2	4/4/2022	mmw	nebuliser

3	4/4/2022	obs 108	2 bp app
4	4/7/2022	MMW	2 bp app
5	4/8/2022	peads ward	mpnitor probe
6	4/11/2022	ot	ot lamp
7	4/13/2022	mow	bp app
8	4/15/2022	icu	ventilator dragger
9	4/15/2022	opthal ot	2 zeiss microscope
10	4/22/2022	obs 108	nst machine
11	4/25/2022	micro lab	bactec machine
12	4/26/2022	nicu	2 monitor
13	4/27/2022	obs 108	bp app
14	4/27/2022	opthal ward	bp app

<u>sr no</u>	<u>DATE</u>	<u>DEPT</u>	<u>EQUIPMENT</u>
1	5/6/2022	Dental opd	suction chair
2	5/11/2022	blood bank	bp app
3	5/16/2022	icu	monitor
4	5/16/2022	obs 108	bp app
5	5/16/2022	radiology casulty	meditronix portable x-ray
6	5/17/2022	msw 205	bp app
7	5/17/2022	nicu	2 monitor
8	5/18/2022	obgy opd	bp app,weight machine
9	5/19/2022	cssd	big autoclave m shah
10	5/21/2022	icu	syringe pump
11	5/21/2022	mmw	2 bp app
12	5/23/2022	nicu	meditrian warmer
13	5/23/2022	icu	nebuliser
14	5/26/2022	fmw	o2 meter
15	5/25/2022	casulty	3 monitor
16	5/27/2022	blood bank	plasma bath
17	5/31/2022	pead ward	bp app
18			

<u>sr no</u>	<u>date</u>	<u>dept</u>	<u>equipment</u>
1	6/1/2022	pnc ward	bp app
2	6/1/2022	obs 108	suction machine
3	6/2/2022	fmw	central o2 meter

4	6/3/2022	obs 108	baby warmer
5	6/7/2022	nicu	bp app
6	6/13/2022	opthal ward	bp app
7	6/13/2022	opthal ot	zeiss microscope
8	6/13/2022	msw	bp app
9	6/13/2022	obs 108	bp app
10	6/17/2022	opthal ot	opthal microscope bulb
11	6/20/2022	mmw	nebuliser
12	6/20/2022	fsw	bp app
13	6/22/2022	dental opd	suction chair
14	6/22/2022	mmw	nebuliser
15	6/23/2022	nicu	bp cuff
16	6/23/2022	nicu	ambu bag
17	6/25/2022	fmw	bp app
18	6/25/2022	casualty	bp app
19	6/25/2022	icu	stress test machine
20	6/30/2022	gyn ward	bp app

sr no	date	dept	equipment
1	2/7/2022	pead ward	bp app/ nebuliser machine
2	2/7/2022	G bb	2 bp app
3	4/7/2022	mow	bp app
4	5/7/2022	ot ent	foot switch (paddle)
5	5/7/2022	obgy opd	2 bp app
6	5/7/2022	radiology	x-ray machine
7	6/7/2022	fsw	o2 meter
8	8/7/2022	mow	needle cutter
9	8/7/2022	casualty	o2 flow meter /small ot lamp /standing lamp
10	11/7/2022	nicu	ventilator bella vista
11	11/7/2022	skin opd	radio cautery dermaindia
12	14/7/2022	casualty	weight machine
13	14/7/2022	msw	bp app
14	14-7-22	Biochemistry	water distillation plant
15	18-7-22	MMW	BP App
16	18-7-22	ICU	Ventilator
17	20-7-22	Micro	Pipate
18	25-7-22	Blood bank	Centrifuge
19	27-7-22	Ophthal OPD	Slit lamp
20	29-7-22	ICU	ventilator

Sr no	Date	Dept	Equipment
1	3/8/2022	Med OPD	6 BP App
2	4/8/2022	MMW	O2 meter cylinder
3	9/8/2022	ICU	Digital BP
4	9/8/2022	Blood bank	Deep Fridge
5	16-8-22	Casualty	Nebulizer
6	16-8-22	Dental	Air roter handpiece
7	26-8-22	Skin OPD	Centrifuge
8	29-8-22	Dialysis	Nipro dialysis
9	30-8-22	Labour	Standing Lamp
10	30-8-22	Casualty	Nebulizer

Sr no	Date	Dept	Equipment
1	1/9/2022	FMW	BP App
2	1/9/2022	FSW	Weight machine
3	1/9/2022	ICU	Ventilator-dragger
4	2/9/2022	ICU	Ventilator-Resmed
5	5/9/2022	Casualty	Monitor
6	6/9/2022	MOW	BP App
7	6/9/2022	Dialysis	Nipo
8	6/9/2022	OT	Stryker Camera
9	6/9/2022	OT	Manman hand key
10	7/9/2022	Pediatric	Monitor
11	7/9/2022	PNC	Baby weight machine
12	8/9/2022	CSDD	ETO
13	8/9/2022	Skin OPD	Phototherapy
14	8/9/2022	Blood Bank	Weight machine
15	12/9/2022	Dental OPD	Dental X ray
16	14-9-22	ICU	Nebulizer
17	14-9-22	ECG	Shiller
18	14-9-22	OBGY	NST
19	14-9-22	OBS-108	Lamp LED

Sr no	Date	Dept	Equipment
1	4/10/2022	Dental OPD	Dental Chair
2	13-10-22	PNC-107	BP app
3	13-10-22	Labour	LED Lamp

4	14-10-22	Dialysis	Dialysis
5	18-10-22	Casualty	Nebuliser
6	18-10-22	PNC-107	BP app
7	18-10-22	Ophthal OPD	Perimeter
8	18-10-22	Ophthal OT	Microscope
9	28-10-22	Ophthal OPD	B scan
10	31-10-22	OBS	NST
11	31-10-22	Ortho OT	Stann Table
12	31-10-22	MSW	BP app

22-Jan

break down
not working
not working
not working
not working
not working
not working
no record seen on screen
bp app brpken
poor conection
o2 meter broken
not working
not working
not working
not working
leakage problm
not working

action taken
repair ,change the baloon ,rubber bag
para issue ,baloon change
change the baloon
conform to company person
baloon, rubber bag, welcrow,change

check out
check out
check out
indent to new

need to change spo2
conform to manman

not working
not working
not working
handle door not working
wire cutting
not working
not working
not working
not working

clean to sand paper

conform to remi person

laryngoscope clean to sand paper
check out & print ok

FEB

break down
heating problm
not working
screen not showing
test fail
not working
not working
not working glass of vision broken
not working
not working
bp cuff tear

action taken
complaint lock
baloon change
conform to company person
caling to company person
caling to company person &request to sending quat

sending to company person
change the welcrow bag

not working
not working
not working
broken
mercury leakage
test tube broken
not working
not working
not working
rotar jam
not working

change the test tube, baloon
change the rubber bag both
check the ventii ,testpass working good
giving to new
inform to engg.&check out

MARCH

break down
abnormal sound is in machine
display problm
not working
broken
weight not showing accurate
O2 not delivered
m/c not working loose wire conection
Flow alarm test not passs
vision not clear
not working
spo2 inner notch broken
table do not up
not working
broken baloon not working
screen not working
display problm
mercury not going up
not working
not working
broken not repairable
not working
not working

action taken
inform to company person
waitnig for repairing from vishnu sir
giving to new
not repairable

inform to company person
complaint lock to company

complain lock to company

change the notch
calling to company person
payment issue so engg not coming to servicing
change the hinjus & baloon
inform to company person
check by vishnu sir &inform sending to quat.
change & clean the mercury
wire soldering

condem
not repairable demand to new

APRIL

BREAKDOWN
not woking
not woking

ACTION TAKEN

mercury leak
MERCURY leak PROBLEM
not working
not working
broken
not working
1 not working
2 foot pads problem
us probe is not working
motor is not working
bp not showing
balloon problem
not working

change the mercury ,balloon, rubber bag
Clean the mercury
change the probe
inform to engg.
change the hinjus
inform to engg.

inform to engg.
waiting for quotation
need to change motor
need to change bp tube
need to change balloon
change the hinjus

may

BREAKDOWN
not working
not working
not working
mercury leak
switch button is not working
not working
bp not showing
not working
not working
due to short crkt pump is burn
not working
not working
broken main cable
not working
not working
not working
not working

ACTION TAKEN
inform to engg
change the velcro bab,rubber bag
need to change setting
clean the mercury
engg came & check
not working
tubing problem
bp app repaired , weight machine broken
waiting for quotation
change the fuse ,main power cord plug
change the rubber bah ,balloon
purchase for new warmer
cutting the wire and also slidering the main cables
change the valve

send to quotation to purchase
change the rubber bag

jun

breakdown
not working
not working
not working

action taken
change the rb & balloon
inform to engg
change the o ring

not working
not working
broken
light problm
not working
not working
not working
wire burn
not working
not working
main cord cut
not working
not working
not working
not working
not working
not working

soldering the wire

repair
inform to engg
change the balloon
mercury leak prolm / not repairable
calling to engg
soldering the wire
change the balloon
calling to engg
change the wire
change the bp cuff
not repairable
change the rb & balloon
change the balloon

change the rb & balloon

july

breakdown
not working / wire broken of nebu
1 test tube broken , 2 not working
not working
broken
not working
allengers control pannel issue
broken
not working
not working
ventii stop frequently while connect to baby
machine created too much noise
broken
not working
water leakage, gasket and pipe damage
Bollon torn
Device failuire
Broken
not working
not moving
Device failuire

action taken
change the balloon

change the rubber bag & balloon
not repairable
not repairable
inform to company person
change the nut of meter

change the O2 meter & given to new
inform to company person
inform to company person
not repairable
change the rubber bag & balloon
replaced with new plant
Ballone changed
software update
given new one
machune is too old ,not repairable
repired
software update

22-Aug

Breakdown

not working
meter not working
not working ,RUSTED
door problem
wire burn due to shortckt
not working
noise problem
not properly, hardness issue
Lamp rod ring broken
wire burn due to shortckt

Action taken

replaced acessories
new meter given
Not repairable
company person came and repaired
soldered wire
repaired
repaired
repaired
repaired
soldered wire

22-Sep

Breakdown

Broken
not working, broken
Device failuar
battery alarm
Not working,
Not working
Not working
Buttun not working
not working
not working
battery and bourd faulty
water leakge problem
SMPS Board Faulty
Not showing wt properly
X ray not shoot
Not working
not charging properly
Speaker properly
not working

Action taken

Hinjes changed
not repairable
repaired
repaired
Sent to company
replace Hinjes
repaired
not repairable
Given new one
repaired
repaired
repaired
Replaced
Glass broken, not repairable
repaired
repaired
not repaired
not repaired
wire broken

22-Oct

Breakdown

scaler not working
Not working
Not working

Action taken

Need servicing
replaced acessories
change power cord

Conductivity test not pass
Not working
Leakage
Not getting started
Light intensity problem
Probe not working
Not working properly
remote problem
Leakage

Repaired
Repaired
Change Accessories
repaired
Replaced Light
sent to company
Speaker problem
Not repaired yet
replaced accessories

22-Nov

repair
repair
repair
repair

repair

repair
repair
repair

not repair

repair

repair

repair
repair

remark

repair

repair

repair

repair

repair

repair

repair

repair

repair

remark

outsource

outsource

outsource

outsource

outsource

repair

repair

outsource

outsource

inhouse

outsource

outsource

inhouse

inhouse

inhouse

inhouse

REMARK

inhouse

inhouse

inhouse
inhouse
inhouse
out source
inhouse
out source

out source
out source
out source
out source
inhouse
inhouse

REMARK
outsourcing
inhouse
inhouse
inhouse
outsourcing
inhouse
outsourcing
inhouse
outsourcing
inhouse
inhouse
outsourcing
inhouse
inhouse

outsourcing
inhouse

remark
inhouse
outsourcing
outsourcing

inhouse
inhouse
inhouse
outsource
inhouse
inhouse
outsource
inhouse
inhouse
outsource
inhouse
inhouse
inhouse
inhouse
inhouse
outsource
inhouse

remark
inhouse
inhouse
inhouse
outsource
inhouse
outsource
inhouse
inhouse
outsource
outsource
outsource
inhouse
inhouse
outsource
inhouse
outsource
outsource
outsource
inhouse
outsource

Remark

inhouse
inhouse
inhouse
Outsource
inhouse
inhouse
inhouse
Outsource
inhouse
inhouse

Remark

Inhouse
Inhouse
Outsource
Inhouse
Outsource
Inhouse
Outsource
Outsource
Outsource
Inhouse
Outsource
Outsource
Outsource
Inhouse
Outsource
Inhouse
Outsource
Outsource
Inhouse

Remark

Outsource
Inhouse
Inhouse

outsource
Inhouse
Inhouse
outsource
Inhouse
outsource
outsource
outsource
Inhouse

Jun-21

Sr no.	Date	Department	Equipment name	Brakedown	Action taken (Inhouse/O utsource)	Remark
1	1/6/2021	Physiotherapy	BP app	ballon torn	Inhouse	
2	3/6/2021	OT	endoscopy	Softwareproblem	Outsource	
3	3/6/2021	NICU	SPO2 probe	sensor faulty	Inhouse	
4	7/6/2021	Ophthal OPD	Vision drum	remot faulty	Inhouse	
5	7/6/2021	Casualty	Monitor	Display not working	Inhouse	
6	10/6/2021	Microbiology	Incubator	Not working	Inhouse	
7	11/6/2021	OT	Warmer	notgetting heat	Inhouse	
8	14-7-21	CSSD	Sealing machine	Not getting heat	Inhouse	
9	15-6-21	NICU	Syringe Pump	Not working	Inhouse	
10	18-6-21	NICU	Syringe Pump	Not working	Inhouse	
11	23-6-21	Microbiology	Autoclave	coil burnt	Inhouse	
12	23-6-21	MMW	BP app	Lekage	Inhouse	
13	23-6-21	FMW	BP app	Not working	Inhouse	
14	23-6-21	Physiotherapy	Ultrasound quality	connection broken	Inhouse	
15	24-6-21	OT	ETCO2 CABLE	Not working	Outsource	
16	25-6-21	PNC	Baby wt machine	Cable broken	Inhouse	
17	29-6-21	Physiology	ECG Cmachine	Battery faulty	new given	
18	30-6-21	OT	Bi-polaer cable	Not working	Inhouse	